



## King's Research Portal

DOI:

[10.1007%2Fs00431-015-2595-4](https://doi.org/10.1007%2Fs00431-015-2595-4)

*Document Version*

Peer reviewed version

[Link to publication record in King's Research Portal](#)

*Citation for published version (APA):*

Shetty, S., Bhat, P., Hickey, A., Peacock, J. L., Anthony D Milner, & Greenough, A. (2016). Proportional assist versus assist control ventilation. *European Journal of Pediatrics*, 175. <https://doi.org/10.1007%2Fs00431-015-2595-4>

### **Citing this paper**

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

### **General rights**

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

### **Take down policy**

If you believe that this document breaches copyright please contact [librarypure@kcl.ac.uk](mailto:librarypure@kcl.ac.uk) providing details, and we will remove access to the work immediately and investigate your claim.

**Proportional assist versus assist control ventilation in premature infants**

**Sandeep Shetty<sup>1</sup>, Prashanth Bhat<sup>2</sup>, Ann Hickey<sup>1</sup>, Janet L Peacock<sup>3,4</sup>,**

**Anthony D Milner<sup>2</sup>, Anne Greenough<sup>2,4</sup>**

**Authors emails:**

Sandeep Shetty: sandeep.shetty1@nhs.net

Prashanth Bhat: prashanth.bhat@kcl.ac.uk

Ann Hickey: annhickey@nhs.net

Janet Peacock: janet.peacock@kcl.ac.uk

Anthony Milner: anthony.milner@kcl.ac.uk

Anne Greenough: anne.greenough@kcl.ac.uk

**Corresponding author:**

Professor Anne Greenough, NICU, 4th Floor Golden Jubilee Wing, King's College Hospital,

Denmark Hill, London SE5 9RS, UK

[Tel:02032993037](tel:02032993037); Fax 02032993617 email: [anne.greenough@kcl.ac.uk](mailto:anne.greenough@kcl.ac.uk)

<sup>1</sup> Neonatal Intensive Care Centre, King's College Hospital, London, UK <sup>2</sup>Division of Asthma, Allergy and Lung Biology, MRC Centre for Allergic Mechanisms in Asthma, King's College London, London, UK <sup>3</sup>Division of Health and Social Care Research, King's College London, London, UK <sup>4</sup>NIHR Biomedical Research Centre at Guy's and St Thomas' NHS Foundation Trust and King's College London, London, UK

## ABSTRACT

During proportional assist ventilation (PAV) the applied pressure is servo-controlled based on continuous input from the infant's breathing. In addition, elastic and resistive unloading can be employed to compensate for the abnormalities in the infant's lung mechanics. The aim of this study was to test the hypothesis that in very prematurely born infants remaining ventilated beyond the first week, PAV compared to assist control ventilation (ACV) would be associated with superior oxygenation. A randomised crossover study was undertaken. Infants were studied for four hours each on PAV and ACV in random order; at the end of each four hour period, the oxygenation index (OI) was calculated. Eight infants, median gestational age of 25 (range 24–33) weeks, were studied at a median of 19 (range 10–105) days. It had been intended to study 18 infants but as all the infants had superior oxygenation on PAV ( $p=0.0039$ ), the study was terminated after recruitment of eight infants. The median inspired oxygen concentration ( $p=0.049$ ), mean airway pressure ( $p=0.012$ ) and OI ( $p=0.012$ ) were all lower on PAV.

**Conclusion:** These results suggest that PAV compared to ACV is advantageous in improving oxygenation for prematurely born infants with evolving or established BPD.

**Key words:** Proportional assist ventilation; assist control ventilation; oxygenation

50    **LIST OF ABBREVIATIONS**

51

52    ACV            Assist control ventilation

53    BPD            Bronchopulmonary dysplasia

54    FiO<sub>2</sub>           Fraction of inspired oxygen concentration

55    MAP            Mean airway pressure

56    OI              Oxygenation index

57    PAV            Proportional assist ventilation

58    PCV            Patient controlled ventilation

59    PEEP           Positive end expiratory pressure

60    PIP            Peak inspiratory pressure

61    SIMV           Synchronised intermittent mandatory ventilation

62

## **What is known**

- During proportional assist ventilation (PAV), the applied pressure is servo controlled throughout each spontaneous breath.
- Elastic and resistive unloading can compensate for the infant's abnormalities in lung mechanics.

## **What is new**

- In a randomised crossover study, infants with evolving/established BPD were studied on PAV and ACV each for four hours.
- The oxygenation index was significantly lower on PAV in all infants studied.

## INTRODUCTION

During proportional assist ventilation (PAV), the applied pressure is servo controlled, based on continuous input from the infant's breathing throughout each spontaneous breath. In addition, the ventilator can provide inflation pressure in phase with the tidal volume change in order to reduce the compliance load (ie, the load due to the stiffness of infant's lungs) and in phase with the flow change to reduce the resistance load (ie, the load due to airflow obstruction), termed elastic and resistive unloading, respectively [6]. Very prematurely born infants developing or with established bronchopulmonary dysplasia (BPD) will have stiff lungs (that is non-compliant) despite a very compliant chest wall, so may be particularly likely to benefit from elastic unloading. PAV has only been assessed in neonates in a few studies. In a previous crossover study, we demonstrated that infants with evolving or established bronchopulmonary dysplasia (BPD) on PAV compared to assist control ventilation (ACV) had better oxygenation indices, a lower work of breathing and better respiratory muscle strength. The infants, however, were only studied on each ventilator mode for one hour [1]. The longest infants have been studied on PAV is four hours [7], but during that study only changes in pulse oximetry results were assessed. During PAV, the applied pressure is servocontrolled throughout each breath, whereas during ACV only the initiation of inflation is synchronised to the start of inspiration, hence we hypothesised that oxygenation would be superior on PAV compared to ACV. During PAV, however, we have demonstrated a trigger delay of 60 msec using an invitro model [4], hence it was important to assess blood gases over a longer period than studied previously [1]. The aim of this study, therefore, was to test the hypothesis that infants with evolving or established BPD would have superior oxygenation index results after four hours on PAV compared to after a similar period on ACV.

## METHODS

A randomised, crossover study was undertaken. Prematurely born infants remaining ventilated after the first week after birth were eligible for entry into the study if they were being supported by ACV. Evolving BPD was defined as ventilator dependence beyond 14 days and established BPD as ventilator dependence beyond 28 days. Infants were ineligible for inclusion in the study if they had a major congenital cardiac abnormality or were receiving a neuromuscular blockade agent. Infants were entered into the study if their parents gave informed written consent. The study was approved by the South East London Research Ethics Committee and King's College Hospital Research Ethics Committee. During the eight hours study period no other changes were made to the infant care than the changes in ventilator mode.

Infants at King's College Hospital NHS Foundation Trust are routinely supported by the SLE 5000. The infants were transferred from the SLE 5000 ventilator to ACV on the Stephanie ventilator using the same ventilator settings (baseline). All infants as per the unit's routine policy were ventilated via shouldered endotracheal tubes which have been shown to have minimal or no leaks [3]. One hour was allowed for stabilisation of the infant on the Stephanie ventilator. A blood gas analysis was then performed and the baseline ventilator settings were noted. During the stabilisation period, the ventilator displayed compliance and resistance settings were noted every ten minutes and the six results meaned. The ventilator calculated the compliance from the inflation pressure (PIP-PEEP) and the resultant tidal volume. The value of the reciprocal of the compliance, elastance, was used to set the level of the elastic unloading. Each infant was then randomised to receive first either PAV or ACV mode for four hours and for the second four hours received the alternative mode. During ACV, the peak inspiratory pressure (PIP), positive end expiratory pressure (PEEP) and the

inflation time were kept the same as at baseline. During PAV, the maximum PIP was set at 5 cm H<sub>2</sub>O above the PIP on ACV. The PEEP level during PAV was the same as at baseline and the PEEP and the inflation time during back up ventilation were the same as at baseline. Whenever cessation of spontaneous breathing occurred for more than five seconds during PAV, mandatory backup inflations were automatically delivered by the ventilator. The backup rate set at 40 breaths per minute was delivered to the infant for 10 seconds with a backup inflation peak pressure of 5 cm H<sub>2</sub>O above the PIP used during ACV. Elastic unloading, which was used only during inspiration, was initially set at 75% of full unloading. Full unloading was the level of unloading which increased the infant's compliance to the expected 'normal', that is, 2.0 mL/cm H<sub>2</sub>O/kg. If after 10 minutes the infant remained stable and no airway pressure waveform abnormalities were observed [4], the unloading was increased to 100%. If airway pressure waveform abnormalities were then noted by looking at the pressure display, the unloading was to be reduced back to 75%. Resistive unloading was not used as, in an vitro model, oscillations in the airway pressure waveforms appeared when the resistive unloading was greater than 100 cm H<sub>2</sub>O/l/sec [4].

The number of desaturations (an oxygen saturation less than 88%) on each mode was noted. An arterial blood sample was obtained at the end of each four hour period, the ventilatory settings were noted and the oxygenation index (OI) was calculated. Respiratory rate, tidal volume and mean airway pressure were obtained from the ventilator. The results from the last five minutes of the four hour study period were averaged. All the infants had continuous oxygen saturation monitoring. During the study, the inspired oxygen concentration was adjusted as necessary to maintain the oxygen saturation level in the range 92-96%.



## **Sample size**

In our previous study, the mean OI in the PAV group was 6.0 (SD  $\pm$  2.4) and in the ACV group was 9.8 (SD  $\pm$  3.7) [1]. The planned sample size was 18 infants to allow detection between the two ventilator modes of a within patient difference of 0.7 SD in the oxygenation index results with 80% power and a two-sided significance of 5%.

## **Analysis**

Differences were assessed for statistical significance using the paired Wilcoxon signed rank test using IBM SPSS statistical software, V.21 (IBM Corporation, USA).

## **RESULTS**

Recruitment to the trial was stopped at eight infants as all the OI results were in favour of PAV (Table 1). The decision was taken by the clinical team and the study statistician in the knowledge that the probability of all 8/8 results in the same direction (hence favouring PAV) if both modes were in fact equally effective was extremely small ( $0.5^8=0.0039$ ).

The median birth weight of the infants was 767 (range 650-1926) gms, gestational age 25.7 (range 24.4-33.5) weeks and postnatal age at measurement 19 (range 10-105) days; seven of the eight infants were male. All the infants had been exposed to antenatal steroids, received postnatal surfactant and were receiving caffeine at the time of study. None were receiving sedation at the time of study or had received postnatal steroids. Their median baseline compliance was 0.4 (range 0.3-1.1) ml/cmH<sub>2</sub>O and resistance was 155 (range 66-252 ) cmH<sub>2</sub>O/l/sec. All infants tolerated 100% elastic unloading throughout the study. The

median  $\text{FiO}_2$  ( $p=0.049$ ), the median mean airway pressure ( $p=0.012$ ) and the median oxygenation index ( $p=0.012$ ) were all lower on PAV compared to ACV (Table 2). There was no significant difference in the median number of desaturation episodes between the two modes.

## DISCUSSION

We have demonstrated that PAV compared to ACV in prematurely born infants ventilated beyond the first week after birth resulted in superior oxygenation index results. Those results are in keeping with those of Schulze et al [7] who compared PAV to SIMV or ACV in infants with evolving BPD. The 22 infants had a median gestational age of 25.6 weeks and were studied at a mean postnatal age of 22.9 days. They found after a four hour period of PAV that despite a lower MAP, the inspired oxygen concentration and pulse oximetry readings were not significantly different between the two groups. All the infants in our study had arterial blood gas measurements and hence we were able to calculate their oxygenation index. We compared PAV to ACV as both modes provide respiratory support for all the infant's breaths. During SIMV, only a preset number of the infant's breaths are supported by the ventilator and hence this might at least partially explain why PAV was superior to SIMV/ACV in the earlier study [7]. We have previously demonstrated that during PAV [1] the inspiratory tidal volume and inflation pressures are closely phase matched and the oesophageal pressure is out of phase as a result of the proportionality. This does not occur in ACV and likely leads to more efficient support during PAV. We did not record whether compliance was improved on PAV compared to ACV, but in our previous paper [1] demonstrated there was an almost significant ( $p=0.05$ ) reduction in thoraco-abdominal asynchrony which could improve oxygenation. In our previous study [1] the respiratory rate was significantly lower during

PAV, but the medians were 54 bpm on PAV and 57 bpm on ACV, so unlikely to be of clinical significance. In this study we did not demonstrate any significant differences in the respiratory rates or delivered tidal volumes. PAV compared to ACV support was not associated with any significant reduction in PaCO<sub>2</sub> but the mean airway pressure was significantly lower during PAV, suggesting the PIP was lower and hence that PAV might have resulted in greater CO<sub>2</sub> clearance.

In a previous four hour cross-over study [7], although the incidence of arterial oxygen desaturations was not significantly different, the desaturations lasted longer when infants were supported by PAV. In that study, however, the infants had a history of frequent apnoeas and arterial oxygen desaturations. None of the infants in this study had been ventilated because of a history of apnoea. In addition, in the previous study [7] a time of 10 seconds was used during which the ventilator software identified cessation of breathing, whereas we used an updated version of the Stephanie software in which a five second period was used during which the ventilator software identified the cessation of breathing. It is likely then that this updated version enabled better support during PAV when the infant was apnoeic. We did not demonstrate any significant difference in the number of desaturations, indeed the infants in both groups experienced very few desaturations.

Our study was terminated before our calculated sample size. We were mindful that all 12 infants in our one hour cross-over study had lower OI results on PAV compared to ACV [1]. Hence, we wished to stop this study as early as possible if all the PAV results were again superior to the ACV results.

In our previous study [1] we also reported that on PAV, the median pressure time product level was significantly lower than on ACV indicating a lower work of breathing, which may reflect more synchronised support by the ventilator throughout inspiration. Neurally adjusted ventilatory assist (NAVA) also applies airway pressure throughout inspiration. During NAVA the pressure applied is proportional to the electrical activity of the diaphragm. In a crossover study of 14 preterm infants [2], asynchrony was significantly lower during 12 hours on NAVA than during 12 hours on pressure regulated, volume controlled ventilation. In addition, amongst ten infants recovering from severe acute respiratory distress syndrome, oxygenation was superior after eight hours of NAVA compared to after eight hours of pressure support ventilation (PSV) [5]. These data [1, 2, 5] and the results currently reported suggest ventilation modes which apply airway pressure in proportion to the infant's respiratory effort throughout inspiration may be superior to those modes in which synchronisation is only at the start of inspiration (ACV), or the start and end of inspiration (pressure support ventilation).

In conclusion, we have demonstrated in a short term cross-over study PAV compared to ACV was associated with significantly superior oxygenation which likely reflects the better synchronisation of the inflation pressure and tidal volume throughout inspiration. We, therefore, feel these data emphasize the need now for a randomised controlled trial.

## ACKNOWLEDGEMENTS

**Contributors:** AG, ADM and PB designed the study. SS undertook all the assessments. JLP advised on the analysis. All authors were involved in the production of the manuscript.

**Funding:** PB was supported by the Charles Wolfson Charitable Trust. The research was supported by the National Institute for Health Research (NIHR) Biomedical Research Centre based at Guy's and St Thomas' NHS Foundation Trust and King's College London. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health. AG is an NIHR Senior Investigator.

**Competing interests:** AG has held grants from various ventilator manufacturers; AG and ADM have received honoraria for giving lectures and advising various ventilator manufacturers.

## REFERENCES

1. Bhat P, Patel DS, Hannam S, Rafferty GF, Peacock JL, Milner AD, Greenough A (2015) Crossover study of proportional assist versus assist control ventilation. *Arch Dis Child Fetal Neonatal Ed* 100:F35-F38
2. Longhini F, Ferrero F, De Luca D, Cosi G, Alemani M, Colombo D, Cammarota G, Berni P, Conti G, Bona G, Della Corte F, Navalesi P (2015) Neurally adjusted ventilatory assist in preterm neonates with acute respiratory failure. *Neonatology* 107:60-67.
3. Hird M, Greenough A, Gamsu HR (1990) Gas trapping during high frequency positive pressure ventilation using conventional ventilators. *Early Hum Dev* 22:51–56
4. Patel DS, Rafferty GF, Hannam S, Lee S, Milner AD, Greenough a (2010) In vitro assessment of proportional assist ventilation. *Arch Dis Child Fetal Neonatal Ed* 95:F331–F337
5. Piastra M, De Luca D, Costa R, Pizza A, De Sanctis R, Marzano L, Biasucci D, Visconti F, Conti G (2014) Neurally adjusted ventilatory assist vs pressure support ventilation in infants recovering from severe acute respiratory distress syndrome: nested study. *J Crit Care* 312:e1-e5.
6. Schulze A (2002) Respiratory mechanical unloading and proportional assist ventilation in infants. *Acta Paediatr Suppl* 437:19–22.
7. Schulze A, Rieger-Fackeldey E, Gerhard T, Claire N, Everett R, Bancalari E (2007) Randomized crossover comparison of proportional assist ventilation and patient-triggered ventilation in extremely low birth weight infants with evolving chronic lung disease. *Neonatology* 92:1–7

284 **Table 1: Oxygenation index results by ventilatory mode**

285 **Individual data are given**

286

INFANT	PAV	ACV
1	8.1	15.2
2	7.0	8.3
3	17.4	21.6
4	12.6	20.3
5	5.2	8.4
6	10.0	14.1
7	5.9	11.3
8	9.9	11.7

287

288

289

290 **Table 2: Comparison of airway pressures and blood gas exchange by ventilator mode**291 **The results are expressed as the median (range)**

	<b>PAV</b>	<b>ACV</b>	<b>P value</b>
FiO <sub>2</sub>	0.48 (0.31-0.65)	0.57 (0.40-0.72)	0.049
PaO <sub>2</sub> (kpa)	5.6 (5.4-8.6)	5.6 (5.0-6.8)	0.327
PCO <sub>2</sub> (kpa)	8.0 (5.5-9.3)	7.2 (5.5-11)	0.889
Oxygenation Index	9.0 (5.2-17.4)	12.9 (8.3-21.6)	0.012
Mean airway pressure (cm H <sub>2</sub> O)	8.5 (6.7 -10.0)	9.5 (8.1 -13)	0.012
Respiratory rate (bpm)	56.6 (47.5-76.7)	57.4 (49.5 -66)	0.401
Tidal volume (mls/kg)	7.5 (3.7-10.0)	5.8 (3.5-10.3)	0.234
Desaturation episodes (n)	1 (0-2)	0 (0-2)	0.429

292

293

294

295

296